EMSL Analytical, Inc.
200 Route 130 North, Cinnaminson, NJ 08077
Phone: (856) 303-2500 Fax: (856) 858-4571 Email: EnvChemistry2@emsl.com

Attn: **Larry Sample**
Prime Example Services, Inc.
123 Main Road
Somewhere, US 01234
Phone: (123) 456-7890
Fax: (987) 654-3210

00/00/2019

The following analytical report covers the analysis performed on samples submitted to EMSL Analytical, Inc. on 00/00/2019. The results are tabulated on the attached data pages for the following client designated project:

Large House

The reference number for these samples is EMSL Order #011910000 Example. Please use this reference when calling about these samples. If you have any questions, please do not hesitate to contact me at (856) 303-2500.

Approved By:

Phillip Worby, Chemistry Laboratory Manager

The test results contained within this report meet the requirements of NELAP and/or the specific certification program that is applicable, unless otherwise noted.
NELAP Certifications: NJ 03036, NY 10872, PA 68-00367, CA ELAP 1877

Report amended 12/00/2019 00:00:00 Replaces initial report from QA TESTING of Amendment

The samples associated with this report were received in good condition unless otherwise noted. This report relates only to those items tested as received by the laboratory. The QC data associated with the sample results meet the recovery and precision requirements established by the NELAP, unless specifically indicated. All results for soil samples are reported on a dry weight basis, unless otherwise noted. This report may not be reproduced except in full and without written approval by EMSL Analytical, Inc.

EMSL Analytical, Inc.
200 Route 130 North, Cinnaminson, NJ 08077
Phone/Fax: (856) 303-2500 / (856) 858-4571
http://www.EMSL.com EnvChemistry2@emsl.com

EMSL Order: 011910000 Example
CustomerID: ABCD12
CustomerPO:
ProjectID:

Attn: **Larry Sample**
Prime Example Services, Inc.
123 Main Road
Somewhere, US 01234

Phone: (123) 456-7890
Fax: (987) 654-3210
Received: 00/00/19 0:00 AM

Project: **Large House**

Analytical Results

Client Sample Description P19-0000-00 ***Collected:*** 00/00/2019 00:00:00 PM ***Lab ID:*** *011910000-0001*

Method	*Parameter*	*Result*	*RL*	*Units*	*Prep Date & Analyst*	*Analysis Date & Analyst*
LC/MS/MS						
EPA 537.1	ADONA	ND	3.6	ng/L	11/27/2019 OM	11/27/19 19:02 OM
4,8-dioxa-3H-perfluorononanoic acid MDL: 0.37 ng/L						
EPA 537.1	N-EtFOSAA	ND	3.6	ng/L	11/27/2019 OM	11/27/19 19:02 OM
N-ethyl perfluorooctanesulfonamidoacetic acid MDL: 0.85 ng/L						
EPA 537.1	9Cl-PF30N	ND	3.6	ng/L	11/27/2019 OM	11/27/19 19:02 OM
9-chlorohexadecafluoro-3-oxanone-1-sulfonic acid MDL: 0.75 ng/L						
EPA 537.1	N-MeFOSAA	ND	3.6	ng/L	11/27/2019 OM	11/27/19 19:02 OM
N-methyl perfluorooctanesulfonamidoacetic acid MDL: 0.67 ng/L						
EPA 537.1	11Cl-PF3C	ND	3.6	ng/L	11/27/2019 OM	11/27/19 19:02 OM
11-chloroeicosafluoro-3-oxaundecane-1-sulfonicacid MDL: 0.83 ng/L						
EPA 537.1	PFBS	ND	3.6	ng/L	11/27/2019 OM	11/27/19 19:02 OM
Perfluorobutanesulfonic acid (PFBS) MDL: 0.68 ng/L						
EPA 537.1	HFPO-DA	ND	3.6	ng/L	11/27/2019 OM	11/27/19 19:02 OM
Hexafluoropropylene oxide dimer acid MDL: 1.1 ng/L						
EPA 537.1	PFDA	ND	3.6	ng/L	11/27/2019 OM	11/27/19 19:02 OM
Perfluorodecanoic acid (PFDA) MDL: 0.67 ng/L						
EPA 537.1	PFDoA	ND	3.6	ng/L	11/27/2019 OM	11/27/19 19:02 OM
Perfluorododecanoic acid (PFDoA) MDL: 0.66 ng/L						
EPA 537.1	PFHpA	ND	3.6	ng/L	11/27/2019 OM	11/27/19 19:02 OM
Perfluoroheptanoic acid (PFHpA) MDL: 0.60 ng/L						
EPA 537.1	PFHxS	ND	3.6	ng/L	11/27/2019 OM	11/27/19 19:02 OM
Perfluorohexanesulfonic acid (PFHxS) MDL: 0.62 ng/L						
EPA 537.1	PFHxA	ND	3.6	ng/L	11/27/2019 OM	11/27/19 19:02 OM
Perfluorohexanoic acid (PFHxA) MDL: 0.60 ng/L						
EPA 537.1	PFNA	ND	3.6	ng/L	11/27/2019 OM	11/27/19 19:02 OM
Perfluorononanoic acid (PFNA) MDL: 0.56 ng/L						
EPA 537.1	PFOS	ND	3.6	ng/L	11/27/2019 OM	11/27/19 19:02 OM
Perfluorooctanesulfonic acid (PFOS) MDL: 0.50 ng/L						
EPA 537.1	PFOA	ND	3.6	ng/L	11/27/2019 OM	11/27/19 19:02 OM
Perfluorooctanoic acid (PFOA) MDL: 1.0 ng/L						
EPA 537.1	PFTeDA	ND	3.6	ng/L	11/27/2019 OM	11/27/19 19:02 OM
Perfluorotetradecanoic acid (PFTA) MDL: 0.55 ng/L						
EPA 537.1	PFTrDA	ND	3.6	ng/L	11/27/2019 OM	11/27/19 19:02 OM
Perfluorotridecanoic acid (PFTrDA) MDL: 0.56 ng/L						
EPA 537.1	PFUdA	ND	3.6	ng/L	11/27/2019 OM	11/27/19 19:02 OM
Perfluoroundecanoic acid (PFUnA) MDL: 0.56 ng/L						

EMSL Analytical, Inc.
200 Route 130 North, Cinnaminson, NJ 08077
Phone/Fax: (856) 303-2500 / (856) 858-4571
http://www.EMSL.com EnvChemistry2@emsl.com

EMSL Order:	011910000 Example
CustomerID:	ABCD12
CustomerPO:	
ProjectID:	

Definitions:

MDL - method detection limit
J - Result was below the reporting limit, but at or above the MDL
ND - indicates that the analyte was not detected at the reporting limit
RL - Reporting Limit (Analytical)
D - Dilution Sample required a dilution which was used to calculate final results