



EMSL Analytical, Inc.

490 Rowley Road Depew, NY 14043
Tel/Fax: (716) 651-0030 / (716) 651-0394
<http://www.EMSL.com> / buffalolab@emsl.com

EMSL Order: 142002938
Customer ID: EMSL14
Customer PO:
Project ID:

Attention: Christopher Goulah
EMSL - BUFFALO
490 Rowley Rd
Depew, NY 14043

Phone: (716) 651-0030
Fax: (716) 651-0394
Collected Date: 07/23/2020
Received Date: 08/04/2020
Analyzed Date: 08/04/2020

Project:

Test Report: Air-O-Cell™ Analysis of Fungal Spores & Particulates by Optical Microscopy (Methods MICRO-SOP-201, ASTM D7391)

Lab Sample Number:	142002938-0001	142002938-0002	142002938-0003		
Client Sample ID:	1	2	3		
Volume (L):	150	150	150		
Sample Location:	Livingroom	Bathroom	Exterior		
Spore Types	Count/M ³	Count/M ³	Count/M ³		
Alternaria (Ulocladium)	-	-	-		
Ascospores	80	40	440		
Aspergillus/Penicillium	-	-	-		
Basidiospores	80	200	250		
Bipolaris++	-	-	-		
Chaetomium	-	-	-		
Cladosporium	20	80	230		
Curvularia	40	80	-		
Epicoccum	20	-	-		
Fusarium	-	-	-		
Ganoderma	-	-	-		
Myxomycetes++	-	-	-		
Pithomyces++	-	-	-		
Rust	-	-	-		
Scopulariopsis/Microascus	-	-	-		
Stachybotrys/Memnoniella	-	40	-		
Unidentifiable Spores	-	-	-		
Zygomycetes	-	-	-		
Triadelphia	20	40	-		
Total Fungi	260	480	920		
Hyphal Fragment	-	-	-		
Insect Fragment	-	-	-		
Pollen	-	-	-		
Analyt. Sensitivity 600x	21	21	21		
Analyt. Sensitivity 300x	7*	7*	7*		
Skin Fragments (1-4)	1	2	1		
Fibrous Particulate (1-4)	1	1	1		
Background (1-5)	1	1	1		

++ Includes other spores with similar morphology; see EMSL's fungal glossary for each specific category.

Christopher Goulah, Microbiology Manager
or other Approved Signatory

No discernable field blank was submitted with this group of samples.

High levels of background particulate can obscure spores and other particulates, leading to underestimation. Background levels of 5 indicate an overloading of background particulates, prohibiting accurate detection and quantification. Present = Spores detected on overloaded samples. Results are not blank corrected unless otherwise noted. The detection limit is equal to one fungal spore, structure, pollen, fiber particle or insect fragment. **** Denotes particles found at 300X. "-" Denotes not detected. Due to method stopping rules, raw counts in excess of 100 are extrapolated based on the percentage analyzed. EMSL maintains liability limited to cost of analysis. Interpretation and use of test results are the responsibility of the client. This report relates only to the samples reported above, and may not be reproduced, except in full, without written approval by EMSL. EMSL bears no responsibility for sample collection activities or analytical method limitations. The report reflects the samples as received. When the information supplied by the customer can affect the validity of the result, it will be noted on the report.
Samples analyzed by EMSL Analytical, Inc. Depew, NY A2LA Accredited Environmental Testing Cert #2845.24

Initial report from: 08/04/2020 14:56:19

For information on the fungi listed in this report, please visit the Resources section at www.emsl.com

Test Report SPVER2-2.10 Printed: 08/04/2020 03:10 PM